



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Icon Clinical Laboratories*, using *Trinity Biotech Premier Hb9210* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Premier Hb9210	Calibrator Lot: 13901, 13902	Column Lot: 13762
Reagent Lot: 12378, 12436, 13657, 12349	Calibrator Assigned Values: 5.50%, 12.50%	

Date of Certification: July 1, 2023

Certification Expires: July 1, 2024

A handwritten signature in black ink, appearing to read "David Sacks".

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Randie R Little PhD".

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read "Sharon Anulf".

SRL director/ supervisor